

# Endotoxin Testing Report

Bacterial Endotoxin Test (LAL)

| SAMPLE INFORMATION |                   | TEST INFORMATION |                                        |
|--------------------|-------------------|------------------|----------------------------------------|
| Sample ID:         | NL-TESA-10-040226 | Method:          | LAL Gel-Clot (sensitivity 0.125 EU/mL) |
| Customer:          | True North Labs   | Status:          | Complete                               |
| Date Collected:    | April 10, 2026    |                  |                                        |
| Date Tested:       | April 14, 2026    |                  |                                        |

## TEST RESULTS

| RESULT        | INTERPRETATION                |
|---------------|-------------------------------|
| < 0.125 EU/mL | Pass — endotoxin not detected |

## NOTES

|                                                                                                          |
|----------------------------------------------------------------------------------------------------------|
| Endotoxin was not detected in Tesamorelin at the validated 1/16 dilution, with a result of <0.125 EU/mL. |
|----------------------------------------------------------------------------------------------------------|

Testides provides analytical testing services and reports laboratory results for the submitted sample only. Testides does not recommend, promote, endorse, certify, or validate the personal or clinical use of research chemicals, nor does it provide medical advice. Any risk assessment language is presented solely as a calculation based on client-provided assumptions (including route of administration and maximum daily dose) and must not be interpreted as a determination of safety, suitability for human use, or regulatory compliance for any finished drug.

